

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAR 25 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130129

Name JEAN ACHEE

ID# 1470

Date 3/6/09

A

Agency WASILLA POLICE DEPT.

Phone # 352-5401

Instrument Location WASILLA POLICE DEPT.

Alco S/N X172942

Target Value .078

High Pressure 300

B

Alco Test Values

.083

1st Alco

.084

2nd Alco

Signature

Jean Achee

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 7th day of May, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

MARCH 06, 2009

OPERATOR'S NAME:

ACHEE, JEAN P

OPERATOR'S NUMBER: 1470

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI:

UOC

O.L. #: 1234567890

DEPT/AGENCY: WASH

CASE/REPORT: 090306

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: X172942

— BREATH ANALYSIS —

.078 ADJUSTED FOR 30.42 in	✓	
ALCO TARGET	.079	23:39
BLANK TEST	.000	00:00
INTERNAL STANDARD	VERIFIED	00:00
ALCO TV 30.42 in	.083	00:01
BLANK TEST	.000	00:01
SUBJECT SAMPLE	.000	00:02
BLANK TEST	.000	00:03
ALCO TV 30.42 in	.084	00:03
BLANK TEST	.000	00:04

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

MARCH 07, 2009 ✓

TIME 00:05

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.42 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () ~ !